

Human Galectin-3 ELISA

Cat. No.: BA1007

Enzyme Immunoassay for the quantitative determination Galectin-3 in human serum and plasma.

Galectin-3 is a member of the galectin family, which are β -galactoside-binding lectins with ≥ 1 evolutionary conserved carbohydrate-recognition domain.¹ Galectin-3 is expressed by various immune cells and is readily secreted by injured and inflammatory cells. Thus, galectin-3 is a novel candidate biomarker for the diagnosis, analysis, and prognosis of various cardiac diseases, including heart failure.²


Galectin-3 has also been linked to kidney disease in clinical studies, and its inhibition appears to improve renal disease.

PRINCIPLE OF GALECTIN-3 ELISA

The microtiter plate is coated with the antibody specifically binding the Galectin-3. The human serum or plasma is incubated in the plate with the capture antibody.

The specimen is washed out and the specifically bound protein is incubated with HRP-labelled detection antibody. Unbound reagent is then washed out. Horseradish peroxidase (HRP) bound in the complex reacts with the chromogenic substrate (TMB) creating the blue colour. The reaction is stopped by addition of STOP solution (H_2SO_4).

The absorbance values are measured at 450 nm (optionally 450/630 nm) and are proportional to the concentration of Galectin-3 in the specimen. The concentration of Galectin-3 in unknown samples is determined from the calibration curve which is created by plotting the absorbance values against the standard concentration values.



Kit Contents

Item	Qty.
Antibody Coated Microtiter Plate	96 wells
Antibody-HRP Conjugate	13 mL
Master Standard (lyophilized)	2 vial
Quality Control A (human serum, lyophilized)	1 vial
Quality Control B (human serum, lyophilized)	1 vial
Dilution Buffer	13 mL
Wash Buffer 15x conc.	50 mL
Substrate Solution	13 mL
STOP Solution	13 mL

MATERIAL REQUIRED BUT NOT SUPPLIED

1. Glassware and test tubes
2. Microtiter plate washer
3. Precision pipettes (various volumes) with tips
4. Orbital shaker
5. Microtiter plate reader capable of measuring absorbance at 450 nm or 450/630 nm with software for data generation

WARNINGS AND PRECAUTIONS

1. For research use only
2. For professional laboratory use
3. The reagents with different lot numbers should not be mixed
4. To prevent cross sample contamination, use disposable labware and pipette tips
5. To protect laboratory stuff, wear protective gloves and protective clothing
6. The substrate solution should remain colourless, keep it protected from light
7. The test should be performed at standard laboratory conditions (temperature $25^\circ\text{C} \pm 2^\circ\text{C}$).

STORAGE CONDITIONS

1. The kit must be stored at $2 - 8^\circ\text{C}$.
2. The opened components can be stored for one week at $2 - 8^\circ\text{C}$.

PREPARATION OF REAGENTS

- Use new pipette tip for pipetting different reagents and samples to prevent cross-contamination.
- All reagents and samples should be allowed to reach the temperature $25^\circ\text{C} \pm 2^\circ\text{C}$.

Preparation of Standards

Reconstitute lyophilized Human Galectin-3 Standard in Dilution Buffer, for the volume information see the Certificate of Analysis. Let it rehydrate for 15 min. The concentration of human Galectin-3 in Master Standard is 4 ng/mL,

Prepare set of Standard solution as follows:

Use the Master Standard to produce a dilution series (as below). Mix each tube thoroughly before the next transfer. The Dilution Buffer serves as Blank.

	Volume of Standard	Dilution Buffer	Concentration
Std1	Standard 4 ng/mL (lyophilized)	500 µL	4 ng/mL
Std2	250 µL of Std1	250 µL	2 ng/mL
Std3	250 µL of Std2	250 µL	1 ng/mL
Std4	250 µL of Std3	250 µL	0.5 ng/mL
Std5	250 µL of Std4	250 µL	0.25 ng/mL
Std6	250 µL of Std5	250 µL	0.125 ng/mL
Blank	-	250 µL	0 ng/mL

Preparation of Quality Control A and B

Reconstitute the lyophilized human serum Quality Controls in deionized/distilled water, for the volume information see the Certificate of Analysis. Let the QCs rehydrate for 15 min and dilute them 1:20 in Dilution Buffer, prior to use, see Preparation of samples.

Preparation of Wash Buffer 1x

Prepare a working solution of Wash Buffer by adding 50 mL of Wash Buffer 15x conc. to 700 mL of deionized/ distilled water (dH₂O). Mix well. Store at 4°C for two weeks or at -20°C for long term storage.

Preparation of samples

Human serum or plasma may be used with this assay. For long-term storage the samples should be frozen at minimum -70°C. Lipemic or haemolytic samples may cause false results.

Recommended dilution of samples is 1:20, i.e., 7 µL of sample + 133 µL of Dilution Buffer for singlets and 14 µL of sample + 266 µL of Dilution Buffer for duplicates.

Do not store the diluted samples.

ASSAY PROCEDURE

1. Prepare the reagents as described in the previous chapter.
2. Pipette 100 µL of set of Standards, Quality Controls, diluted Samples and Dilution Buffer = Blank into each well. Incubate for **1 hour** at 25°C ±2°C, shaking at 300 rpm.
3. Wash the wells 3-times with 1x Wash Buffer (350 µL/well). When finished, tap the plate against the paper towel to remove the liquid completely.
4. Pipette 100 µL of HRP-labelled Antibody Conjugate into each well. Incubate for **1 hour** at 25°C ±2°C, shaking at 300 rpm.
5. Wash the wells as described in point 3.
6. Pipette 100 µL Substrate solution, incubate for **10 min**, at 25°C ±2°C. Avoid exposure to the light during this step.
7. Pipette 100 µL of STOP solution.
8. Read the signal at 450 or 450/630 nm within 15 min.

PERFORMANCE CHARACTERISTICS

Samples used in the tests were diluted 1:20 as recommended and assayed. The results are multiplied by the dilution factor.

1. Sensitivity

The limit of detection, defined as a concentration of human Galectin-3 giving absorbance higher than absorbance of blank + 3 standard deviations, is better than 0.03 ng/mL of sample.

2. Precision

Intra-assay

Sample	Mean (ng/mL)	SD	CV (%)
1	8.6	0.4	5
2	10.4	0.5	5

Inter-assay (Run – to – run)

Sample	Mean (ng/mL)	SD	CV (%)
1	4.9	0.4	9
2	9.0	0.3	3

3. Accuracy

Dilution linearity


Sample	Dilution	Measured concentration (ng/mL)	Expected concentration (ng/mL)	Yield (%)
1		8.5	-	-
	2x	4.4	4.3	104
	4x	2.5	2.1	116
	8x	1.3	1.1	119
2		9.9	-	-
	2x	4.9	5.0	99
	4x	2.7	2.5	108
	8x	1.4	1.2	113

Spiking Recovery

Sample	Spike (ng/mL)	Measured concentration (ng/mL)	Expected concentration (ng/mL)	Yield (%)
1	-	9.0	-	-
	20	30.2	29.0	104
	10	18.6	19.0	98
	4	11.6	13.0	90

Typical standard curve

The standard curve needs to be measured in every test. Most of the microplate reader can automatically calculate the analyte concentration using 4-parameter algorithm or alternative functions to fit the standard points properly. The concentrations need to be multiplied by the dilution factor, either automatically by reader or manually.



RESOURCES

¹ Dong R, Zhang M, Hu Q, Zheng S, Soh A, Zheng Y, Yuan H. Galectin-3 as a novel biomarker for disease diagnosis and a target for therapy (Review). *Int J Mol Med*. 2018 Feb;41(2):599-614. doi: 10.3892/ijmm.2017.3311. Epub 2017 Dec 5. PMID: 29207027; PMCID: PMC5752178.

² Hara, A.; Niwa, M.; Kanayama, T.; Noguchi, K.; Niwa, A.; Matsuo, M.; Kuroda, T.; Hatano, Y.; Okada, H.; Tomita, H. Galectin-3: A Potential Prognostic and Diagnostic Marker for Heart Disease and Detection of Early Stage Pathology. *Biomolecules* 2020, 10, 1277. <https://doi.org/10.3390/biom10091277>